

Calibration Certificate & Function Test

Gerät / Device / Appareillage / Apparecchio

Instrument	TL-1D
Version number	v2.0
Serial number	A17010123
Measuring range	-20...70 °C
Accuracy	±0.3 K

Allgemein / General / Général / Generale

ROTRONIC AG certifies that this instrument meets the specifications of our datasheet. It has been calibrated and corresponds to the test requirements of **ISO9001-2008**. The instrument mentioned above was calibrated by comparison of the temperature values to the HygroClip HC2-S factory transfer standards. The traceability of the factory transfer standards is given by calibration at an **ISO 17025 (SCS-065)** accredited calibration laboratory. The **SCS-065** calibration laboratory is traceable to the national standard at the **Federal Institute of Metrology (METAS)**.

Referenz / Reference / Référence / Referenza

Parameter	Instrument	Serial number	Calibration date
Temperature	HC2-S	20048862	09.05.2016

Kalibrierung / Calibration / Étalonnage / Calibratura

Temperature

No.	Reference [°C]	Reading [°C]	Deviation [°C]
1	25.4	25.2	-0.2

N/A

No.	Reference []	Reading []	Deviation []

N/A

No.	Reference []	Reading []	Deviation []

Funktionstest / Function test /

Test de fonctionnement / Test funzionali

Display	OK
Keys	OK
Power	OK
Communication	OK

Datum / Date / Date / Date 05.12.2016
Prüfer / Inspector / Vérificateur / Verificatore 15035
This document has been automatically created